

Attestation of Conformity



Korea

Manufacturer **CERAGEM MEDISYS Inc.**
16, Jeongja 1-gil, Seonggeo-eup,
Seobuk-gu, Cheonan-si
Chungcheongnam-do 331-833, Republic of KOREA

Item Identification **CERA-CHEK 1Code Blood Glucose Monitoring System**
(Model No. G400)

Above mentioned product manufactured by CERAGEM MEDISYS Inc.
meets the Minimum System Accuracy Performance Criteria,
defined in the clause 6.3.3 of ISO 15197:2013.

This attestation is released based on the Technical File Assessment Result as they
submitted a Change Notification regarding above mentioned product. The detailed
assessment results are listed in

Test Report No. 74940486
EC-Certificate No. V1 14 05 73112 014

Issued Place: Seoul
Medical Health Services
of TÜV SÜD Korea Ltd.

Subsidiary of
TÜV SÜD Product Service GmbH

Date: 2014-12-30

Dr. Takeo Ushiki

Manager of
Medical Health Services